

Hydrophobic Foldable IOL USFDA

AMI No. 21.301

PL NO. 82932128

Specification

1. It should be Preloaded Hydrophobic acrylic with covalently bonded Yellow Chromophore IOL with automated CO2 powered delivery system.
2. The Delivery speed should be up to 3mm /sec.
3. It should be single use disposable.
4. It should have Aspheric -0.20um, Refractive Index - 1.55 , L - Shaped Modified haptics
5. The optic size should be 6.0mm overall length 13.0mm.
6. It should have Compatible incision size not more than 2.2mm
7. It should have UV and High energy Blue light filtering
8. A- Constant should be - 119.1
9. The Diopter Range should be from + 6.0D to + 30.0D with 0.5D incremental.
10. It should be BIS / USFDA / European CE certified.
11. The company should give undertaking to replace the power within 48 hours as and when required by the hospital.

An
29/05/26
डॉ. अनिता कुमारी
Dr. Anita Kumari
वरिष्ठ. मं. चि. अक्षि. / नेत्र रोग विशेषज्ञ
Sr. DMO/ Ophtha
उ. र. मंडल चिकित्सालय, अम्बाला छावनी
N.R. Divn. Hosp., Ambala Cantt